

**GAS-FIRED COMPRESSOR ENGINE  
PORTABLE EMISSION TEST REPORT  
EQM GATHERING OPCO, LLC  
MCINTOSH COMPRESSOR STATION  
ENGINE 4 - SOURCE ID 104  
FACILITY ID 805715  
PERMIT PA-63-01003A  
UNION TOWNSHIP, WASHINGTON COUNTY, PA**

Test Date: August 27, 2019

Report Date: September 4, 2019



**Prepared by:**

Equitrans Midstream  
544 Copley Road  
Weston, WV 26452  
724-852-7417

## TABLE OF CONTENTS

	<u>Page</u>
1 TEST RESULTS SUMMARY	1
2 INTRODUCTION & CONTACT INFORMATION	2
3 TEST DATE AND PERSONNEL INFORMATION	3
4 SOURCE DESCRIPTION AND PROCESS DATA	3
4.1 Source Description	3
4.2 Process Data	3
5 TESTING PROCEDURES	4
5.1 Field Work	4
5.2 Determination of Gaseous Emissions	4
5.2.1 Sampling System Setup	
5.2.2 Measurement by ASTM D6522	
6 CONCLUSION	6

Table: Table 1- Emission Test Results, Engine 4, Caterpillar 3616TALE, 5,350 BHP

Section I: Certificate of Accuracy

Section II: Results Summary

Section III: Instrument Calibrations

Section IV: Site Data

Section V: Run Data

Section VI: Gas Analysis

Section VII: Calibration Gas Certificates

**GAS-FIRED COMPRESSOR ENGINE  
 PORTABLE EMISSION TEST REPORT  
 EQM GATHERING OPCO, LLC  
 MCINTOSH COMPRESSOR STATION  
 ENGINE 4 - SOURCE ID 104  
 UNION TOWNSHIP, WASHINGTON COUNTY, PA**

**1. TEST RESULTS SUMMARY**

<b>Pennsylvania Department of Environmental Protection (PADEP)</b>			
<b>Permit # PA-63-01003A</b>			
<b>Caterpillar G3616LE – Source ID 104</b>			
<b>Pollutant</b>	<b>Average Result*</b>	<b>Permit Limits</b>	<b>Compliant/Non-Compliant</b>
Oxides of Nitrogen	2.59 lb./hr	-	-
	11.32 tpy	-	-
	0.24 g/bhp-hr	0.30	Compliant
	18.82 ppmvd @15% O <sub>2</sub>	-	-
Carbon Monoxide	0.24 lb./hr	-	-
	1.07 tpy	-	-
	0.02 g/bhp-hr	0.17	Compliant
	2.91 ppmvd @15% O <sub>2</sub>	-	-

\*tpy based on 8,760 operating hours per year

## **2. INTRODUCTION**

Testing was conducted in accordance with the Equitrans Portable Performance Test Protocol. The source tested was one (1) Caterpillar Gas-fired compressor engine. The engine has a rating of 5,350 brake horsepower (BHP) at 1000 revolutions per minute (rpm). Testing was conducted for oxygen (O<sub>2</sub>), oxides of nitrogen (NO<sub>x</sub> – NO and NO<sub>2</sub>) and carbon monoxide (CO). Testing was conducted in accordance with United States Environmental Protection Agency (USEPA) Methods 18 and 19 as detailed in 40 Code of Federal Regulations (CFR), Part 60, Appendix A and ASTM D6522. Testing was conducted to demonstrate compliance with the emission limits of PADEP Permit # PA-63-01003A.

## **CONTACT INFORMATION**

Mr. Willoughby Hill  
Equitrans Midstream  
544 Copley Road  
Weston, West Virginia 26452  
(724) 852-7417 – Office  
(304) 550-2453 – Mobile  
[whill@equitransmidstream.com](mailto:whill@equitransmidstream.com)

Mr. Carl Ent  
Equitrans Midstream  
317 East Roy Furman Hwy  
Waynesburg, PA 15370  
(412) 328-6248 – Mobile  
[cent@equitransmidstream.com](mailto:cent@equitransmidstream.com)

### 3. TEST DATE AND PERSONNEL INFORMATION

The testing occurred on August 27, 2019 at the McIntosh Compressor Station and was performed according to USEPA Methodology. The following table details the contact information regarding this test program:

<b>Organization</b>	<b>Personnel</b>	<b>Responsibility</b>
Equitrans	Mr. Carl Ent	Operator, RMs 18, 19, and ASTMD6522
Equitrans	Mr. Willoughby Hill	Supervisor

### 4. SOURCE DESCRIPTION AND PROCESS DATA

#### 4.1 Source Description

The internal combustion engine is used to drive a compressor to pressurize natural gas for delivery to market. The engine tested at this location was (1) 5,350 brake horsepower Caterpillar Model G3616LE reciprocating internal combustion natural gas-fired compressor engine.

#### 4.2 Process Data

Equitrans recorded the following process parameters during the time of the testing:

- Engine speed
- Intake Manifold Pressure (IMAP)
- Ignition Timing
- Fuel Flow Rate
- Engine Load
- BHP (calculated)

Appendix F contains the process data.

## **5. TESTING PROCEDURES**

Testing was conducted in accordance with USEPA, Title 40, Code of Federal Regulations (CFR), Part 60, Appendix A, Methods 18 and 19. Additionally, testing was conducted using ASTM D6522.

### **5.1 Field Work**

All field and Excel data sheets for the test procedures are included in the Appendices.

### **5.2 Determination of Gaseous Emissions**

#### **5.2.1 Sampling System Setup**

The sampling system consisted of a sample probe, heated sample line, sample conditioner, and emission analyzer. A stainless steel “tee” in the sample probe facilitated system calibration. A heated Teflon sample line delivered the sample gas to the sample conditioning system and instrumentation system, which was located at ground level. The emission analyzer is complete with a self-contained pump that draws the sample from the exhaust stack at approximately 1.1 L/min for analysis. Prior to the analyzer, a sample conditioner reduces the temperature of the sample in order to condense and remove water vapor from the sample stream. The sample then passes through a filter to remove any contaminants in the sample stream. A peristaltic pump is attached to the outlet of the conditioner in order to reject any moisture condensed from the system. A calibration manifold is available in order to inject calibration gases for pre- and post- analysis instrument verification.




Figure 6-1 Sampling Schematic

### 5.2.2 Measurement by ASTM D6522

Prior to testing, the sample system was allowed to achieve steady state operating temperature. Calibration gases of NO, NO<sub>2</sub>, CO, and O<sub>2</sub> were introduced to the sample system inlet at values representative of exhaust stack concentrations. Instrument response was recorded and was within 5% of prescribed calibration gas values. After pre-run verification, the sample probe was inserted into the exhaust system and exhaust gas was drawn through the sample system, into the analyzer for a period of 1-hour. Measurements were taken at 30 second intervals and averaged to achieve test results. Following the test run, calibration gases were sent to the sample probe once more for post-run verification and instrument response. The instrument response was within the 5%

tolerance of prescribed calibration value. Pre- and post- run responses were used to correct the test run average for system bias and drift.

## **6. CONCLUSION**

An emission evaluation test program has been conducted on one 5,350 bhp compressor engine at the McIntosh Compressor Station located in Washington County, Pennsylvania. Testing was conducted for O<sub>2</sub>, CO, and NO<sub>x</sub>. Test results are considered to be representative of the emission rates at the prevailing operating conditions.




## Section I.

### Certificate of Accuracy

### Certification of Data Accuracy

I, the undersigned, hereby certify that, based on information and belief formed after reasonable inquiry, all information contained in the attached McIntosh Engine #4 Source ID:104 - Portable Emission Test Report, and any supporting documents appended hereto, is true, accurate, and complete.


**Signature**  \_\_\_\_\_ **Date** 9/17/2019  
Responsible of Authorized Representative

**Name** William Sybert (WJS) **Title** Vice President Engineering

**Telephone #** 412-395-7041

### Certification of Compliance Status

I, the undersigned, hereby certify that, based on the finding(s) of the test(s) conducted on the date(s), beginning August 27, 2019 and ending August 27, 2019, show that the facility McIntosh Compressor Station Engine #4 Source ID: 104, is in compliance  X, or is not in compliance \_\_\_\_\_, with Rule/Condition # Section E.I. #001 of Permit PA-63-01003A.

**Signature**  \_\_\_\_\_ **Date** 9/17/2019  
Responsible of Authorized Representative

**Name** William Sybert (WJS) **Title** Vice President Engineering

**Telephone #** 412-395-7041

**E-Mail Address** wsybert@equitransmidstream.com

## Section II.

### Results Summary

Equitrans Midstream

Emission Summary for

McIntosh

Engine #

4

Station:  
Engine #:  
HP:  
Data Collected by:

McIntosh  
4  
5350  
C.Ent

Manfct:  
Model:  
Rated Speed:  
Date:

Caterpillar  
3616 A4  
1000  
8/27/2019

Engine Serial Number:  
Engine Hrs

ZZY00735  
6,924

Date	8/27/2019	Emission Summary	Measured	Permit levels
Time	11:08:08 AM			
Ambient Conditions		NOx TPY	11.32	
Rel Hum %	94.0%	NOx Lb/Hr	2.59	
Ambient Air Temp	68.0	NOx g/BHP/Hr	0.24	0.30 <b>Compliant</b>
Barometric Pressure ("Hg)	29.95			
Engine Operating Conditions		CO TPY	1.07	
RPM	999.6	CO Lb/hr	0.24	
% Load	90%	CO g/BHP/Hr	0.02	0.17 <b>Compliant</b>
Ignition Timing	17.9			
Air Man. Temp.	108.4			
AMP (Trailer Inst)	33.3			
Air Man Pressure ("Hg)	33.28			
BTU Content HHV	1042.0			
BTU Content LHV	948.2			
Fuel Flow (MillionBTU/hr)	37.37			
BHP	4825.7			
Exhaust Gas Conditions (Corrected for drift)		CC515961	NO2	
O2 %	11.03	CC1205	NO	
CO ppmvd	4.87	CC515961	O2	
NO ppmvd	22.19	Zero Gas	UHP Nitrogen	
NO2 ppmvd	9.30			
NOx ppmvd	31.49			
CO ppmvd @ 15%	2.91			
NO ppmvd @ 15%	13.26			
NO2 ppmvd @ 15%	5.56			
NOx ppmvd @ 15%	18.82			
Exhaust Emission				
NOx Lb/Hr	2.59			
CO Lb/Hr	0.24			
NOx g/BHP/ Hr	0.24			
CO g/BHP/ Hr	0.02			
NOx TPY	11.32			
CO TPY	1.07			

Section III.  
Calibrations

	Cal Gas	Min ppm	Max ppm	Max Error ppm	% error	Allowable %	Bottle s/n
Stability Check	<b>NO2</b>	<b>101.30</b>	100.10	101.50	<b>1.20</b>	<b>0.39</b>	3 or 5
	<b>NO</b>	<b>199.90</b>	198.00	204.00	<b>4.10</b>	<b>0.68</b>	3 or 5
	<b>O2</b>	<b>15.05</b>	15.02	15.08	<b>0.03</b>	<b>0.07</b>	1% O2
	<b>CO</b>	<b>200.70</b>	191.00	197.00	<b>9.70</b>	<b>1.61</b>	3 or 5

CC507258  
 SG91017577ALD1  
 CC507258  
 CC6889

Date/ Time	sec Runtime	ppm NO2	ppm NO	ppm NOx	% O2	ppm CO
3/20/2019 3:28:52 PM	0	39.8	4	43.8	15.02	0
3/20/2019 3:29:22 PM	30	49.8	3	52.8	15.02	0
3/20/2019 3:29:52 PM	60	49.9	3	52.9	15.02	0
3/20/2019 3:30:22 PM	90	49.9	3	52.9	15.03	0
3/20/2019 3:30:52 PM	120	50.0	3	53.0	15.03	0
3/20/2019 3:31:22 PM	150	49.9	3	52.9	15.02	0
3/20/2019 3:31:52 PM	180	49.9	3	52.9	15.02	0
3/20/2019 3:32:22 PM	210	50.0	3	53.0	15.03	0
3/20/2019 3:32:52 PM	240	50.0	3	53.0	15.02	0
3/20/2019 3:33:22 PM	270	50.0	3	53.0	15.02	0
3/20/2019 3:33:52 PM	300	50.0	3	53.0	15.03	0
3/20/2019 3:34:22 PM	330	50.0	3	53.0	15.03	0
3/20/2019 3:34:52 PM	360	50.0	3	53.0	15.02	0
3/20/2019 3:35:22 PM	390	50.0	3	53.0	15.03	0
3/20/2019 3:35:52 PM	420	50.0	3	53.0	15.02	0
3/20/2019 3:36:22 PM	450	94.9	1	95.9	15.08	0
3/20/2019 3:36:52 PM	480	99.3	4	103.3	15.05	0
3/20/2019 3:37:22 PM	510	99.9	4	103.9	15.05	0
3/20/2019 3:37:52 PM	540	100.1	5	105.1	15.05	0
3/20/2019 3:38:22 PM	570	100.4	5	105.4	15.05	0
3/20/2019 3:38:52 PM	600	100.6	5	105.6	15.05	0
3/20/2019 3:39:22 PM	630	100.6	5	105.6	15.05	0
3/20/2019 3:39:52 PM	660	100.8	5	105.8	15.05	0
3/20/2019 3:40:22 PM	690	100.9	5	105.9	15.05	0
3/20/2019 3:40:52 PM	720	100.8	5	105.8	15.05	0
3/20/2019 3:41:22 PM	750	101.0	5	106.0	15.05	0
3/20/2019 3:41:52 PM	780	101.0	5	106.0	15.05	0
3/20/2019 3:42:22 PM	810	101.2	5	106.2	15.05	0
3/20/2019 3:42:52 PM	840	101.1	6	107.1	15.05	0
3/20/2019 3:43:22 PM	870	101.1	5	106.1	15.05	0
3/20/2019 3:43:52 PM	900	101.1	6	107.1	15.05	0
3/20/2019 3:44:22 PM	930	101.2	6	107.2	15.05	0
3/20/2019 3:44:52 PM	960	101.2	6	107.2	15.05	0
3/20/2019 3:45:22 PM	990	101.2	6	107.2	15.05	0
3/20/2019 3:45:52 PM	1020	101.3	6	107.3	15.05	0
3/20/2019 3:46:22 PM	1050	101.2	6	107.2	15.05	0
3/20/2019 3:46:52 PM	1080	101.2	6	107.2	15.05	0
3/20/2019 3:47:22 PM	1110	101.3	6	107.3	15.05	0
3/20/2019 3:47:52 PM	1140	101.3	6	107.3	15.05	0
3/20/2019 3:48:22 PM	1170	101.3	6	107.3	15.05	0
3/20/2019 3:48:52 PM	1200	101.3	6	107.3	15.05	0
3/20/2019 3:49:22 PM	1230	101.4	6	107.4	15.05	0
3/20/2019 3:49:52 PM	1260	101.3	6	107.3	15.05	0
3/20/2019 3:50:22 PM	1290	101.3	6	107.3	15.05	0
3/20/2019 3:50:52 PM	1320	101.3	6	107.3	15.05	0
3/20/2019 3:51:22 PM	1350	101.4	6	107.4	15.05	0
3/20/2019 3:51:52 PM	1380	101.4	6	107.4	15.05	0
3/20/2019 3:52:22 PM	1410	101.5	6	107.5	15.05	0

3/20/2019 3:52:52 PM	1440	101.4	6	107.4	15.05	0
3/20/2019 3:53:22 PM	1470	101.5	6	107.5	15.05	0
3/20/2019 3:53:52 PM	1500	101.4	6	107.4	15.05	0
3/20/2019 3:54:22 PM	1530	74.0	7	81.0	8.62	0
3/20/2019 3:54:52 PM	1560	3.2	18	21.2	0.07	0
3/20/2019 3:55:22 PM	1590	2.0	18	20.0	0.03	0
3/20/2019 3:55:52 PM	1620	1.5	17	18.5	0.00	0
3/20/2019 3:56:22 PM	1650	1.3	17	18.3	0.00	0
3/20/2019 3:56:52 PM	1680	1.1	17	18.1	0.00	0
3/20/2019 3:57:22 PM	1710	1.0	17	18.0	0.00	0
3/20/2019 3:57:52 PM	1740	0.9	17	17.9	0.00	0
3/20/2019 3:58:22 PM	1770	0.8	17	17.8	0.00	0
3/20/2019 3:58:52 PM	1800	0.7	16	16.7	0.00	0
3/20/2019 3:59:22 PM	1830	0.7	16	16.7	0.00	0
3/20/2019 3:59:52 PM	1860	0.6	16	16.6	0.00	0
3/20/2019 4:00:22 PM	1890	0.6	16	16.6	0.00	0
3/20/2019 4:00:52 PM	1920	0.6	16	16.6	0.00	0
3/20/2019 4:01:22 PM	1950	0.6	16	16.6	0.00	0
3/20/2019 4:01:52 PM	1980	0.6	16	16.6	0.00	0
3/20/2019 4:02:22 PM	2010	0.5	16	16.5	0.00	0
3/20/2019 4:02:52 PM	2040	0.6	16	16.6	0.00	0
3/20/2019 4:03:22 PM	2070	0.5	16	16.5	0.00	0
3/20/2019 4:03:52 PM	2100	0.5	16	16.5	0.00	0
3/20/2019 4:04:22 PM	2130	0.5	16	16.5	0.00	0
3/20/2019 4:04:52 PM	2160	0.5	16	16.5	0.00	0
3/20/2019 4:05:22 PM	2190	0.2	3	3.2	0.00	122
3/20/2019 4:05:52 PM	2220	0.2	1	1.2	0.00	190
3/20/2019 4:06:22 PM	2250	0.2	1	1.2	0.00	192
3/20/2019 4:06:52 PM	2280	0.1	1	1.1	0.00	193
3/20/2019 4:07:22 PM	2310	0.1	1	1.1	0.00	194
3/20/2019 4:07:52 PM	2340	0.1	1	1.1	0.00	195
3/20/2019 4:08:22 PM	2370	0.1	1	1.1	0.00	193
3/20/2019 4:08:52 PM	2400	0.1	1	1.1	0.00	194
3/20/2019 4:09:22 PM	2430	0.1	1	1.1	0.00	196
3/20/2019 4:09:52 PM	2460	0.1	0	0.1	0.00	191
3/20/2019 4:10:22 PM	2490	0.1	0	0.1	0.00	194
3/20/2019 4:10:52 PM	2520	0.1	0	0.1	0.00	196
3/20/2019 4:11:22 PM	2550	0.1	0	0.1	0.00	194
3/20/2019 4:11:52 PM	2580	0.1	0	0.1	0.00	194
3/20/2019 4:12:22 PM	2610	0.1	0	0.1	0.00	195
3/20/2019 4:12:52 PM	2640	0.1	0	0.1	0.00	195
3/20/2019 4:13:22 PM	2670	0.1	0	0.1	0.00	195
3/20/2019 4:13:52 PM	2700	0.1	0	0.1	0.00	194
3/20/2019 4:14:22 PM	2730	0.1	0	0.1	0.00	194
3/20/2019 4:14:52 PM	2760	0.1	0	0.1	0.00	196
3/20/2019 4:15:22 PM	2790	0.1	0	0.1	0.00	194
3/20/2019 4:15:52 PM	2820	0.1	0	0.1	0.00	195
3/20/2019 4:16:22 PM	2850	0.1	0	0.1	0.00	195
3/20/2019 4:16:52 PM	2880	0.1	0	0.1	0.00	195
3/20/2019 4:17:22 PM	2910	0.1	0	0.1	0.00	192
3/20/2019 4:17:52 PM	2940	0.1	0	0.1	0.00	195
3/20/2019 4:18:22 PM	2970	0.1	0	0.1	0.00	194
3/20/2019 4:18:52 PM	3000	0.1	0	0.1	0.00	194
3/20/2019 4:19:22 PM	3030	0.1	0	0.1	0.00	193
3/20/2019 4:19:52 PM	3060	0.1	0	0.1	0.00	195
3/20/2019 4:20:22 PM	3090	0.1	0	0.1	0.00	195
3/20/2019 4:20:52 PM	3120	0.1	0	0.1	0.00	193
3/20/2019 4:21:22 PM	3150	0.1	0	0.1	0.00	197
3/20/2019 4:21:52 PM	3180	0.1	0	0.1	0.00	195
3/20/2019 4:22:22 PM	3210	0.1	0	0.1	0.00	195
3/20/2019 4:22:52 PM	3240	0.1	0	0.1	0.00	195
3/20/2019 4:23:22 PM	3270	0.1	0	0.1	0.00	200

3/20/2019 4:23:52 PM	3300	0.1	0	0.1	0.00	195
3/20/2019 4:24:22 PM	3330	0.1	0	0.1	0.00	197
3/20/2019 4:24:52 PM	3360	0.1	0	0.1	0.42	90
3/20/2019 4:25:22 PM	3390	0.1	0	0.1	0.00	15
3/20/2019 4:25:52 PM	3420	0.1	0	0.1	0.00	11
3/20/2019 4:26:22 PM	3450	0.1	0	0.1	0.00	11
3/20/2019 4:26:52 PM	3480	0.1	0	0.1	0.00	11
3/20/2019 4:27:22 PM	3510	0.1	0	0.1	0.00	10
3/20/2019 4:27:52 PM	3540	0.1	0	0.1	0.00	11
3/20/2019 4:28:22 PM	3570	0.1	0	0.1	0.00	9
3/20/2019 4:28:52 PM	3600	0.1	0	0.1	0.00	10
3/20/2019 4:29:22 PM	3630	0.1	0	0.1	0.00	10
3/20/2019 4:29:52 PM	3660	0.1	0	0.1	0.00	9
3/20/2019 4:30:22 PM	3690	0.1	0	0.1	0.00	8
3/20/2019 4:30:52 PM	3720	0.1	0	0.1	0.00	9
3/20/2019 4:31:22 PM	3750	0.1	0	0.1	0.00	10
3/20/2019 4:31:52 PM	3780	0.1	0	0.1	0.00	9
3/20/2019 4:32:22 PM	3810	0.1	0	0.1	0.00	9
3/20/2019 4:32:52 PM	3840	0.1	0	0.1	0.00	10
3/20/2019 4:33:22 PM	3870	0.1	0	0.1	0.00	10
3/20/2019 4:33:52 PM	3900	0.1	0	0.1	0.00	8
3/20/2019 4:34:22 PM	3930	5.7	134	139.7	0.00	5
3/20/2019 4:34:52 PM	3960	1.5	188	189.5	0.00	0
3/20/2019 4:35:22 PM	3990	1.2	194	195.2	0.00	0
3/20/2019 4:35:52 PM	4020	1.1	196	197.1	0.00	0
3/20/2019 4:36:22 PM	4050	1.0	197	198.0	0.00	0
3/20/2019 4:36:52 PM	4080	0.9	198	198.9	0.00	0
3/20/2019 4:37:22 PM	4110	0.9	199	199.9	0.00	0
3/20/2019 4:37:52 PM	4140	0.8	199	199.8	0.00	0
3/20/2019 4:38:22 PM	4170	0.8	200	200.8	0.00	0
3/20/2019 4:38:52 PM	4200	0.7	200	200.7	0.00	0
3/20/2019 4:39:22 PM	4230	0.7	200	200.7	0.00	0
3/20/2019 4:39:52 PM	4260	0.6	201	201.6	0.00	0
3/20/2019 4:40:22 PM	4290	0.7	201	201.7	0.00	0
3/20/2019 4:40:52 PM	4320	0.6	201	201.6	0.00	0
3/20/2019 4:41:22 PM	4350	0.6	202	202.6	0.00	0
3/20/2019 4:41:52 PM	4380	0.5	202	202.5	0.00	0
3/20/2019 4:42:22 PM	4410	0.5	202	202.5	0.00	0
3/20/2019 4:42:52 PM	4440	0.6	202	202.6	0.00	0
3/20/2019 4:43:22 PM	4470	0.5	202	202.5	0.00	0
3/20/2019 4:43:52 PM	4500	0.5	202	202.5	0.00	0
3/20/2019 4:44:22 PM	4530	0.5	202	202.5	0.00	0
3/20/2019 4:44:52 PM	4560	0.5	202	202.5	0.00	0
3/20/2019 4:45:22 PM	4590	0.5	203	203.5	0.00	0
3/20/2019 4:45:52 PM	4620	0.4	203	203.4	0.00	0
3/20/2019 4:46:22 PM	4650	0.4	203	203.4	0.00	0
3/20/2019 4:46:52 PM	4680	0.5	203	203.5	0.00	0
3/20/2019 4:47:22 PM	4710	0.4	203	203.4	0.00	0
3/20/2019 4:47:52 PM	4740	0.4	203	203.4	0.00	0
3/20/2019 4:48:22 PM	4770	0.4	203	203.4	0.00	0
3/20/2019 4:48:52 PM	4800	0.4	203	203.4	0.00	0
3/20/2019 4:49:22 PM	4830	0.4	203	203.4	0.00	0
3/20/2019 4:49:52 PM	4860	0.4	203	203.4	0.00	0
3/20/2019 4:50:22 PM	4890	0.4	203	203.4	0.00	0
3/20/2019 4:50:52 PM	4920	0.4	204	204.4	0.00	0
3/20/2019 4:51:22 PM	4950	0.3	204	204.3	0.00	0
3/20/2019 4:51:52 PM	4980	0.4	204	204.4	0.00	0
3/20/2019 4:52:22 PM	5010	0.3	204	204.3	0.00	0
3/20/2019 4:52:52 PM	5040	0.3	204	204.3	0.00	0
3/20/2019 4:53:22 PM	5070	6.4	203	209.4	0.00	0
3/20/2019 4:53:52 PM	5100	0.5	29	29.5	3.80	0
3/20/2019 4:54:22 PM	5130	0.0	9	9.0	3.85	0



3/20/2019 4:54:52 PM	5160	0.0	6	6.0	3.86	0
3/20/2019 4:55:22 PM	5190	0.0	5	5.0	3.86	0
3/20/2019 4:55:52 PM	5220	0.0	5	5.0	3.87	0
3/20/2019 4:56:22 PM	5250	0.0	4	4.0	3.87	0
3/20/2019 4:56:52 PM	5280	0.0	4	4.0	3.87	0
3/20/2019 4:57:22 PM	5310	0.0	3	3.0	3.87	0
3/20/2019 4:57:52 PM	5340	0.0	3	3.0	3.87	0
3/20/2019 4:58:22 PM	5370	0.0	3	3.0	3.87	0
3/20/2019 4:58:52 PM	5400	0.0	3	3.0	3.88	0
3/20/2019 4:59:22 PM	5430	0.0	3	3.0	3.88	0
3/20/2019 4:59:52 PM	5460	0.0	3	3.0	3.88	0
3/20/2019 5:00:22 PM	5490	0.0	2	2.0	3.88	0
3/20/2019 5:00:52 PM	5520	0.0	2	2.0	3.88	0
3/20/2019 5:01:22 PM	5550	0.0	2	2.0	3.88	0
3/20/2019 5:01:52 PM	5580	0.0	2	2.0	3.88	0
3/20/2019 5:02:22 PM	5610	0.0	2	2.0	3.88	0
3/20/2019 5:02:52 PM	5640	0.0	2	2.0	3.88	0
3/20/2019 5:03:22 PM	5670	0.0	2	2.0	3.88	0
3/20/2019 5:03:52 PM	5700	0.0	2	2.0	3.88	0
3/20/2019 5:04:22 PM	5730	0.0	2	2.0	3.88	0
3/20/2019 5:04:52 PM	5760	0.0	2	2.0	3.88	0
3/20/2019 5:05:22 PM	5790	0.0	2	2.0	3.88	0

**NO2**

3/20/2019 3:37:52 PM	540	100.1	5	105.1	15.05	0
3/20/2019 3:38:22 PM	570	100.4	5	105.4	15.05	0
3/20/2019 3:38:52 PM	600	100.6	5	105.6	15.05	0
3/20/2019 3:39:22 PM	630	100.6	5	105.6	15.05	0
3/20/2019 3:39:52 PM	660	100.8	5	105.8	15.05	0
3/20/2019 3:40:22 PM	690	100.9	5	105.9	15.05	0
3/20/2019 3:40:52 PM	720	100.8	5	105.8	15.05	0
3/20/2019 3:41:22 PM	750	101.0	5	106.0	15.05	0
3/20/2019 3:41:52 PM	780	101.0	5	106.0	15.05	0
3/20/2019 3:42:22 PM	810	101.2	5	106.2	15.05	0
3/20/2019 3:42:52 PM	840	101.1	6	107.1	15.05	0
3/20/2019 3:43:22 PM	870	101.1	5	106.1	15.05	0
3/20/2019 3:43:52 PM	900	101.1	6	107.1	15.05	0
3/20/2019 3:44:22 PM	930	101.2	6	107.2	15.05	0
3/20/2019 3:44:52 PM	960	101.2	6	107.2	15.05	0
3/20/2019 3:45:22 PM	990	101.2	6	107.2	15.05	0
3/20/2019 3:45:52 PM	1020	101.3	6	107.3	15.05	0
3/20/2019 3:46:22 PM	1050	101.2	6	107.2	15.05	0
3/20/2019 3:46:52 PM	1080	101.2	6	107.2	15.05	0
3/20/2019 3:47:22 PM	1110	101.3	6	107.3	15.05	0
3/20/2019 3:47:52 PM	1140	101.3	6	107.3	15.05	0
3/20/2019 3:48:22 PM	1170	101.3	6	107.3	15.05	0
3/20/2019 3:48:52 PM	1200	101.3	6	107.3	15.05	0
3/20/2019 3:49:22 PM	1230	101.4	6	107.4	15.05	0
3/20/2019 3:49:52 PM	1260	101.3	6	107.3	15.05	0
3/20/2019 3:50:22 PM	1290	101.3	6	107.3	15.05	0
3/20/2019 3:50:52 PM	1320	101.3	6	107.3	15.05	0
3/20/2019 3:51:22 PM	1350	101.4	6	107.4	15.05	0
3/20/2019 3:51:52 PM	1380	101.4	6	107.4	15.05	0
3/20/2019 3:52:22 PM	1410	101.5	6	107.5	15.05	0

**NO**

3/20/2019 4:36:52 PM	4080	0.9	198	198.9	0.00	0
3/20/2019 4:37:22 PM	4110	0.9	199	199.9	0.00	0
3/20/2019 4:37:52 PM	4140	0.8	199	199.8	0.00	0
3/20/2019 4:38:22 PM	4170	0.8	200	200.8	0.00	0
3/20/2019 4:38:52 PM	4200	0.7	200	200.7	0.00	0
3/20/2019 4:39:22 PM	4230	0.7	200	200.7	0.00	0

3/20/2019 4:39:52 PM	4260	0.6	201	201.6	0.00	0
3/20/2019 4:40:22 PM	4290	0.7	201	201.7	0.00	0
3/20/2019 4:40:52 PM	4320	0.6	201	201.6	0.00	0
3/20/2019 4:41:22 PM	4350	0.6	202	202.6	0.00	0
3/20/2019 4:41:52 PM	4380	0.5	202	202.5	0.00	0
3/20/2019 4:42:22 PM	4410	0.5	202	202.5	0.00	0
3/20/2019 4:42:52 PM	4440	0.6	202	202.6	0.00	0
3/20/2019 4:43:22 PM	4470	0.5	202	202.5	0.00	0
3/20/2019 4:43:52 PM	4500	0.5	202	202.5	0.00	0
3/20/2019 4:44:22 PM	4530	0.5	202	202.5	0.00	0
3/20/2019 4:44:52 PM	4560	0.5	202	202.5	0.00	0
3/20/2019 4:45:22 PM	4590	0.5	203	203.5	0.00	0
3/20/2019 4:45:52 PM	4620	0.4	203	203.4	0.00	0
3/20/2019 4:46:22 PM	4650	0.4	203	203.4	0.00	0
3/20/2019 4:46:52 PM	4680	0.5	203	203.5	0.00	0
3/20/2019 4:47:22 PM	4710	0.4	203	203.4	0.00	0
3/20/2019 4:47:52 PM	4740	0.4	203	203.4	0.00	0
3/20/2019 4:48:22 PM	4770	0.4	203	203.4	0.00	0
3/20/2019 4:48:52 PM	4800	0.4	203	203.4	0.00	0
3/20/2019 4:49:22 PM	4830	0.4	203	203.4	0.00	0
3/20/2019 4:49:52 PM	4860	0.4	203	203.4	0.00	0
3/20/2019 4:50:22 PM	4890	0.4	203	203.4	0.00	0
3/20/2019 4:50:52 PM	4920	0.4	204	204.4	0.00	0
3/20/2019 4:51:22 PM	4950	0.3	204	204.3	0.00	0

**O2**

3/20/2019 3:28:52 PM	0	39.8	4	43.8	15.02	0
3/20/2019 3:29:22 PM	30	49.8	3	52.8	15.02	0
3/20/2019 3:29:52 PM	60	49.9	3	52.9	15.02	0
3/20/2019 3:30:22 PM	90	49.9	3	52.9	15.03	0
3/20/2019 3:30:52 PM	120	50.0	3	53.0	15.03	0
3/20/2019 3:31:22 PM	150	49.9	3	52.9	15.02	0
3/20/2019 3:31:52 PM	180	49.9	3	52.9	15.02	0
3/20/2019 3:32:22 PM	210	50.0	3	53.0	15.03	0
3/20/2019 3:32:52 PM	240	50.0	3	53.0	15.02	0
3/20/2019 3:33:22 PM	270	50.0	3	53.0	15.02	0
3/20/2019 3:33:52 PM	300	50.0	3	53.0	15.03	0
3/20/2019 3:34:22 PM	330	50.0	3	53.0	15.03	0
3/20/2019 3:34:52 PM	360	50.0	3	53.0	15.02	0
3/20/2019 3:35:22 PM	390	50.0	3	53.0	15.03	0
3/20/2019 3:35:52 PM	420	50.0	3	53.0	15.02	0
3/20/2019 3:36:22 PM	450	94.9	1	95.9	15.08	0
3/20/2019 3:36:52 PM	480	99.3	4	103.3	15.05	0
3/20/2019 3:37:22 PM	510	99.9	4	103.9	15.05	0
3/20/2019 3:37:52 PM	540	100.1	5	105.1	15.05	0
3/20/2019 3:38:22 PM	570	100.4	5	105.4	15.05	0
3/20/2019 3:38:52 PM	600	100.6	5	105.6	15.05	0
3/20/2019 3:39:22 PM	630	100.6	5	105.6	15.05	0
3/20/2019 3:39:52 PM	660	100.8	5	105.8	15.05	0
3/20/2019 3:40:22 PM	690	100.9	5	105.9	15.05	0
3/20/2019 3:40:52 PM	720	100.8	5	105.8	15.05	0
3/20/2019 3:41:22 PM	750	101.0	5	106.0	15.05	0
3/20/2019 3:41:52 PM	780	101.0	5	106.0	15.05	0
3/20/2019 3:42:22 PM	810	101.2	5	106.2	15.05	0
3/20/2019 3:42:52 PM	840	101.1	6	107.1	15.05	0
3/20/2019 3:43:22 PM	870	101.1	5	106.1	15.05	0

**CO**

3/20/2019 4:07:22 PM	2310	0.1	1	1.1	0.00	194
3/20/2019 4:07:52 PM	2340	0.1	1	1.1	0.00	195

3/20/2019 4:08:22 PM	2370	0.1	1	1.1	0.00	193
3/20/2019 4:08:52 PM	2400	0.1	1	1.1	0.00	194
3/20/2019 4:09:22 PM	2430	0.1	1	1.1	0.00	196
3/20/2019 4:09:52 PM	2460	0.1	0	0.1	0.00	191
3/20/2019 4:10:22 PM	2490	0.1	0	0.1	0.00	194
3/20/2019 4:10:52 PM	2520	0.1	0	0.1	0.00	196
3/20/2019 4:11:22 PM	2550	0.1	0	0.1	0.00	194
3/20/2019 4:11:52 PM	2580	0.1	0	0.1	0.00	194
3/20/2019 4:12:22 PM	2610	0.1	0	0.1	0.00	195
3/20/2019 4:12:52 PM	2640	0.1	0	0.1	0.00	195
3/20/2019 4:13:22 PM	2670	0.1	0	0.1	0.00	195
3/20/2019 4:13:52 PM	2700	0.1	0	0.1	0.00	194
3/20/2019 4:14:22 PM	2730	0.1	0	0.1	0.00	194
3/20/2019 4:14:52 PM	2760	0.1	0	0.1	0.00	196
3/20/2019 4:15:22 PM	2790	0.1	0	0.1	0.00	194
3/20/2019 4:15:52 PM	2820	0.1	0	0.1	0.00	195
3/20/2019 4:16:22 PM	2850	0.1	0	0.1	0.00	195
3/20/2019 4:16:52 PM	2880	0.1	0	0.1	0.00	195
3/20/2019 4:17:22 PM	2910	0.1	0	0.1	0.00	192
3/20/2019 4:17:52 PM	2940	0.1	0	0.1	0.00	195
3/20/2019 4:18:22 PM	2970	0.1	0	0.1	0.00	194
3/20/2019 4:18:52 PM	3000	0.1	0	0.1	0.00	194
3/20/2019 4:19:22 PM	3030	0.1	0	0.1	0.00	193
3/20/2019 4:19:52 PM	3060	0.1	0	0.1	0.00	195
3/20/2019 4:20:22 PM	3090	0.1	0	0.1	0.00	195
3/20/2019 4:20:52 PM	3120	0.1	0	0.1	0.00	193
3/20/2019 4:21:22 PM	3150	0.1	0	0.1	0.00	197
3/20/2019 4:21:52 PM	3180	0.1	0	0.1	0.00	195

	Cal Gas	Min ppm	Max ppm	Max Error ppm	% error	Allowable %	Bottle s/n
Linearity	NO2	51.15	50.00	50.20	1.15	0.75	3 or 5
		95.73	94.60	97.10	1.37	0.48	3 or 5
	NO	50.76	52.00	52.00	1.24	0.81	3 or 5
		197.60	198.00	200.00	2.40	0.40	3 or 5
	O2 in %	15.04	14.99	15.00	0.05	0.11	1% O2
		3.99	3.88	3.90	0.11	0.92	1% O2
	CO	51.19	51.00	52.00	0.81	0.53	3 or 5
		203.30	206.00	206.00	2.70	0.44	3 or 5

Date / time	sec Runtim	ppm NO2	ppm NO	ppm NOx	% O2	ppm CO
8/8/2019 2:41:49 PM	0	0.1	4	4.1	21.31	0
8/8/2019 2:42:19 PM	30	42.6	0	42.6	15.27	1
8/8/2019 2:42:49 PM	60	47.7	0	47.7	15.03	0
8/8/2019 2:43:19 PM	90	48.7	1	49.7	15.01	0
8/8/2019 2:43:49 PM	120	49.2	1	50.2	15.01	0
8/8/2019 2:44:19 PM	150	49.3	1	50.3	15.02	0
8/8/2019 2:44:49 PM	180	49.6	1	50.6	15.00	0
8/8/2019 2:45:19 PM	210	49.8	1	50.8	15.00	0
8/8/2019 2:45:49 PM	240	49.7	1	50.7	15.00	0
8/8/2019 2:46:19 PM	270	49.8	2	51.8	14.98	0
8/8/2019 2:46:49 PM	300	49.9	2	51.9	14.99	0
8/8/2019 2:47:19 PM	330	50.0	2	52.0	15.00	0
8/8/2019 2:47:49 PM	360	50.0	2	52.0	15.00	0
8/8/2019 2:48:19 PM	390	50.0	2	52.0	14.99	0
8/8/2019 2:48:49 PM	420	50.1	2	52.1	14.99	0
8/8/2019 2:49:19 PM	450	50.0	2	52.0	15.00	0
8/8/2019 2:49:49 PM	480	50.1	2	52.1	14.99	0
8/8/2019 2:50:19 PM	510	50.0	2	52.0	14.99	0
8/8/2019 2:50:49 PM	540	50.1	2	52.1	15.00	0
8/8/2019 2:51:19 PM	570	50.1	2	52.1	15.00	0
8/8/2019 2:51:49 PM	600	50.2	2	52.2	14.99	0
8/8/2019 2:52:19 PM	630	50.2	2	52.2	15.00	0
8/8/2019 2:52:49 PM	660	50.2	2	52.2	14.99	0
8/8/2019 2:53:19 PM	690	50.2	2	52.2	15.00	0
8/8/2019 2:53:49 PM	720	50.3	2	52.3	14.99	0
8/8/2019 2:54:19 PM	750	28.3	4	32.3	18.09	0
8/8/2019 2:54:49 PM	780	85.8	1	86.8	14.99	1
8/8/2019 2:55:19 PM	810	94.6	2	96.6	15.00	0
8/8/2019 2:55:49 PM	840	95.1	2	97.1	15.00	0
8/8/2019 2:56:19 PM	870	95.4	3	98.4	15.01	0
8/8/2019 2:56:49 PM	900	95.8	3	98.8	15.01	0
8/8/2019 2:57:19 PM	930	96.0	3	99.0	15.01	0
8/8/2019 2:57:49 PM	960	96.3	3	99.3	15.01	0
8/8/2019 2:58:19 PM	990	96.5	4	100.5	15.01	0
8/8/2019 2:58:49 PM	1020	96.8	4	100.8	15.01	0
8/8/2019 2:59:19 PM	1050	96.9	4	100.9	15.01	0
8/8/2019 2:59:49 PM	1080	97.1	4	101.1	15.01	0
8/8/2019 3:00:19 PM	1110	97.2	4	101.2	15.01	0
8/8/2019 3:00:49 PM	1140	97.2	4	101.2	15.01	0
8/8/2019 3:01:19 PM	1170	97.2	4	101.2	15.01	0
8/8/2019 3:01:49 PM	1200	46.9	16	62.9	4.20	0

8/8/2019 3:02:19 PM	1230	9.7	39	48.7	0.10	0
8/8/2019 3:02:49 PM	1260	3.7	51	54.7	0.05	0
8/8/2019 3:03:19 PM	1290	2.6	52	54.6	0.03	0
8/8/2019 3:03:49 PM	1320	2.0	52	54.0	0.02	0
8/8/2019 3:04:19 PM	1350	1.8	52	53.8	0.00	0
8/8/2019 3:04:49 PM	1380	1.5	52	53.5	0.00	0
8/8/2019 3:05:19 PM	1410	1.3	52	53.3	0.00	0
8/8/2019 3:05:49 PM	1440	1.2	52	53.2	0.00	0
8/8/2019 3:06:19 PM	1470	1.1	52	53.1	0.00	0
8/8/2019 3:06:49 PM	1500	1.0	52	53.0	0.00	0
8/8/2019 3:07:19 PM	1530	1.0	52	53.0	0.00	0
8/8/2019 3:07:49 PM	1560	0.9	52	52.9	0.00	0
8/8/2019 3:08:19 PM	1590	0.8	52	52.8	0.00	0
8/8/2019 3:08:49 PM	1620	0.8	52	52.8	0.00	0
8/8/2019 3:09:19 PM	1650	0.7	53	53.7	0.00	0
8/8/2019 3:09:49 PM	1680	0.9	32	32.9	10.16	0
8/8/2019 3:10:19 PM	1710	40.0	85	125.0	10.06	1
8/8/2019 3:10:49 PM	1740	1.6	9	10.6	20.79	0
8/8/2019 3:11:19 PM	1770	0.9	5	5.9	21.04	0
8/8/2019 3:11:49 PM	1800	4.7	133	137.7	1.76	0
8/8/2019 3:12:19 PM	1830	1.5	189	190.5	0.08	0
8/8/2019 3:12:49 PM	1860	1.3	193	194.3	0.04	0
8/8/2019 3:13:19 PM	1890	1.1	195	196.1	0.02	0
8/8/2019 3:13:49 PM	1920	0.9	196	196.9	0.00	0
8/8/2019 3:14:19 PM	1950	0.8	197	197.8	0.00	0
8/8/2019 3:14:49 PM	1980	0.6	198	198.6	0.00	0
8/8/2019 3:15:19 PM	2010	0.6	198	198.6	0.00	0
8/8/2019 3:15:49 PM	2040	0.5	198	198.5	0.00	0
8/8/2019 3:16:19 PM	2070	0.4	199	199.4	0.00	0
8/8/2019 3:16:49 PM	2100	0.4	199	199.4	0.00	0
8/8/2019 3:17:19 PM	2130	0.3	199	199.3	0.00	0
8/8/2019 3:17:49 PM	2160	0.3	199	199.3	0.00	0
8/8/2019 3:18:19 PM	2190	0.3	199	199.3	0.00	0
8/8/2019 3:18:49 PM	2220	0.2	200	200.2	0.00	0
8/8/2019 3:19:19 PM	2250	0.2	200	200.2	0.00	0
8/8/2019 3:19:49 PM	2280	0.6	144	144.6	1.11	0
8/8/2019 3:20:19 PM	2310	0.0	18	18.0	3.88	0
8/8/2019 3:20:49 PM	2340	0.0	9	9.0	3.88	0
8/8/2019 3:21:19 PM	2370	0.0	7	7.0	3.88	0
8/8/2019 3:21:49 PM	2400	0.0	6	6.0	3.88	0
8/8/2019 3:22:19 PM	2430	0.0	5	5.0	3.88	0
8/8/2019 3:22:49 PM	2460	0.0	5	5.0	3.88	0
8/8/2019 3:23:19 PM	2490	0.0	4	4.0	3.88	0
8/8/2019 3:23:49 PM	2520	0.0	4	4.0	3.88	0
8/8/2019 3:24:19 PM	2550	0.0	3	3.0	3.88	0
8/8/2019 3:24:49 PM	2580	0.0	3	3.0	3.88	0
8/8/2019 3:25:19 PM	2610	0.0	3	3.0	3.88	0
8/8/2019 3:25:49 PM	2640	0.0	3	3.0	3.89	0
8/8/2019 3:26:19 PM	2670	0.1	3	3.1	3.88	0
8/8/2019 3:26:49 PM	2700	0.0	2	2.0	3.88	0
8/8/2019 3:27:19 PM	2730	0.0	2	2.0	3.88	0
8/8/2019 3:27:49 PM	2760	0.0	2	2.0	3.88	0
8/8/2019 3:28:19 PM	2790	0.1	2	2.1	3.88	0
8/8/2019 3:28:49 PM	2820	0.1	2	2.1	3.90	0
8/8/2019 3:29:19 PM	2850	0.0	2	2.0	0.04	48
8/8/2019 3:29:49 PM	2880	0.1	1	1.1	0.00	51
8/8/2019 3:30:19 PM	2910	0.1	1	1.1	0.00	51
8/8/2019 3:30:49 PM	2940	0.1	1	1.1	0.00	51
8/8/2019 3:31:19 PM	2970	0.1	1	1.1	0.00	51
8/8/2019 3:31:49 PM	3000	0.0	1	1.0	0.00	51
8/8/2019 3:32:19 PM	3030	0.1	1	1.1	0.00	51
8/8/2019 3:32:49 PM	3060	0.1	1	1.1	0.00	51
8/8/2019 3:33:19 PM	3090	0.1	1	1.1	0.00	51

8/8/2019 3:33:49 PM	3120	0.1	1	1.1	0.00	52
8/8/2019 3:34:19 PM	3150	0.1	1	1.1	0.00	51
8/8/2019 3:34:49 PM	3180	0.1	1	1.1	0.00	51
8/8/2019 3:35:19 PM	3210	0.1	1	1.1	0.00	52
8/8/2019 3:35:49 PM	3240	0.1	1	1.1	0.00	51
8/8/2019 3:36:19 PM	3270	0.1	1	1.1	0.00	51
8/8/2019 3:36:49 PM	3300	9.5	0	9.5	0.33	31
8/8/2019 3:37:19 PM	3330	0.4	1	1.4	0.02	174
8/8/2019 3:37:49 PM	3360	0.2	1	1.2	0.00	199
8/8/2019 3:38:19 PM	3390	0.2	1	1.2	0.00	203
8/8/2019 3:38:49 PM	3420	0.1	1	1.1	0.00	204
8/8/2019 3:39:19 PM	3450	0.1	1	1.1	0.00	204
8/8/2019 3:39:49 PM	3480	0.1	1	1.1	0.00	205
8/8/2019 3:40:19 PM	3510	0.1	1	1.1	0.00	205
8/8/2019 3:40:49 PM	3540	0.1	1	1.1	0.00	205
8/8/2019 3:41:19 PM	3570	0.1	1	1.1	0.00	206
8/8/2019 3:41:49 PM	3600	0.1	1	1.1	0.00	206
8/8/2019 3:42:19 PM	3630	0.1	1	1.1	0.00	206
8/8/2019 3:42:49 PM	3660	0.1	0	0.1	0.00	206
8/8/2019 3:43:19 PM	3690	0.1	1	1.1	0.00	206
8/8/2019 3:43:49 PM	3720	0.1	0	0.1	0.00	206
8/8/2019 3:44:19 PM	3750	0.1	1	1.1	0.00	206
8/8/2019 3:44:49 PM	3780	0.1	1	1.1	0.00	206
8/8/2019 3:45:19 PM	3810	0.1	1	1.1	0.00	206
8/8/2019 3:45:49 PM	3840	0.1	0	0.1	0.00	206
8/8/2019 3:46:19 PM	3870	0.1	0	0.1	0.00	206
8/8/2019 3:46:49 PM	3900	0.1	1	1.1	0.00	206
8/8/2019 3:47:19 PM	3930	0.1	0	0.1	0.00	206
8/8/2019 3:47:49 PM	3960	0.1	1	1.1	0.00	206
8/8/2019 3:48:19 PM	3990	0.1	0	0.1	0.00	206

NO2 Line 1

8/8/2019 2:47:49 PM	360	50.0	2	52.0	15.00	0
8/8/2019 2:48:19 PM	390	50.0	2	52.0	14.99	0
8/8/2019 2:48:49 PM	420	50.1	2	52.1	14.99	0
8/8/2019 2:49:19 PM	450	50.0	2	52.0	15.00	0
8/8/2019 2:49:49 PM	480	50.1	2	52.1	14.99	0
8/8/2019 2:50:19 PM	510	50.0	2	52.0	14.99	0
8/8/2019 2:50:49 PM	540	50.1	2	52.1	15.00	0
8/8/2019 2:51:19 PM	570	50.1	2	52.1	15.00	0
8/8/2019 2:51:49 PM	600	50.2	2	52.2	14.99	0
8/8/2019 2:52:19 PM	630	50.2	2	52.2	15.00	0

NO2 Line 2

8/8/2019 2:55:19 PM	810	94.6	2	96.6	15.00	0
8/8/2019 2:55:49 PM	840	95.1	2	97.1	15.00	0
8/8/2019 2:56:19 PM	870	95.4	3	98.4	15.01	0
8/8/2019 2:56:49 PM	900	95.8	3	98.8	15.01	0
8/8/2019 2:57:19 PM	930	96.0	3	99.0	15.01	0
8/8/2019 2:57:49 PM	960	96.3	3	99.3	15.01	0
8/8/2019 2:58:19 PM	990	96.5	4	100.5	15.01	0
8/8/2019 2:58:49 PM	1020	96.8	4	100.8	15.01	0
8/8/2019 2:59:19 PM	1050	96.9	4	100.9	15.01	0
8/8/2019 2:59:49 PM	1080	97.1	4	101.1	15.01	0

NO Line 1

8/8/2019 3:03:19 PM	1290	2.6	52	54.6	0.03	0
8/8/2019 3:03:49 PM	1320	2.0	52	54.0	0.02	0
8/8/2019 3:04:19 PM	1350	1.8	52	53.8	0.00	0
8/8/2019 3:04:49 PM	1380	1.5	52	53.5	0.00	0
8/8/2019 3:05:19 PM	1410	1.3	52	53.3	0.00	0
8/8/2019 3:05:49 PM	1440	1.2	52	53.2	0.00	0
8/8/2019 3:06:19 PM	1470	1.1	52	53.1	0.00	0

	8/8/2019 3:06:49 PM	1500	1.0	52	53.0	0.00	0
	8/8/2019 3:07:19 PM	1530	1.0	52	53.0	0.00	0
NO Line 2							
	8/8/2019 3:14:49 PM	1980	0.6	198	198.6	0.00	0
	8/8/2019 3:15:19 PM	2010	0.6	198	198.6	0.00	0
	8/8/2019 3:15:49 PM	2040	0.5	198	198.5	0.00	0
	8/8/2019 3:16:19 PM	2070	0.4	199	199.4	0.00	0
	8/8/2019 3:16:49 PM	2100	0.4	199	199.4	0.00	0
	8/8/2019 3:17:19 PM	2130	0.3	199	199.3	0.00	0
	8/8/2019 3:17:49 PM	2160	0.3	199	199.3	0.00	0
	8/8/2019 3:18:19 PM	2190	0.3	199	199.3	0.00	0
	8/8/2019 3:18:49 PM	2220	0.2	200	200.2	0.00	0
	8/8/2019 3:19:19 PM	2250	0.2	200	200.2	0.00	0
O2 Line 1							
	8/8/2019 2:47:49 PM	360	50.0	2	52.0	15.00	0
	8/8/2019 2:48:19 PM	390	50.0	2	52.0	14.99	0
	8/8/2019 2:48:49 PM	420	50.1	2	52.1	14.99	0
	8/8/2019 2:49:19 PM	450	50.0	2	52.0	15.00	0
	8/8/2019 2:49:49 PM	480	50.1	2	52.1	14.99	0
	8/8/2019 2:50:19 PM	510	50.0	2	52.0	14.99	0
	8/8/2019 2:50:49 PM	540	50.1	2	52.1	15.00	0
	8/8/2019 2:51:19 PM	570	50.1	2	52.1	15.00	0
	8/8/2019 2:51:49 PM	600	50.2	2	52.2	14.99	0
	8/8/2019 2:52:19 PM	630	50.2	2	52.2	15.00	0
O2 Line 2							
	8/8/2019 3:24:19 PM	2550	0.0	3	3.0	3.88	0
	8/8/2019 3:24:49 PM	2580	0.0	3	3.0	3.88	0
	8/8/2019 3:25:19 PM	2610	0.0	3	3.0	3.88	0
	8/8/2019 3:25:49 PM	2640	0.0	3	3.0	3.89	0
	8/8/2019 3:26:19 PM	2670	0.1	3	3.1	3.88	0
	8/8/2019 3:26:49 PM	2700	0.0	2	2.0	3.88	0
	8/8/2019 3:27:19 PM	2730	0.0	2	2.0	3.88	0
	8/8/2019 3:27:49 PM	2760	0.0	2	2.0	3.88	0
	8/8/2019 3:28:19 PM	2790	0.1	2	2.1	3.88	0
	8/8/2019 3:28:49 PM	2820	0.1	2	2.1	3.90	0
CO Line 1							
	8/8/2019 3:29:49 PM	2880	0.1	1	1.1	0.00	51
	8/8/2019 3:30:19 PM	2910	0.1	1	1.1	0.00	51
	8/8/2019 3:30:49 PM	2940	0.1	1	1.1	0.00	51
	8/8/2019 3:31:19 PM	2970	0.1	1	1.1	0.00	51
	8/8/2019 3:31:49 PM	3000	0.0	1	1.0	0.00	51
	8/8/2019 3:32:19 PM	3030	0.1	1	1.1	0.00	51
	8/8/2019 3:32:49 PM	3060	0.1	1	1.1	0.00	51
	8/8/2019 3:33:19 PM	3090	0.1	1	1.1	0.00	51
	8/8/2019 3:33:49 PM	3120	0.1	1	1.1	0.00	52
	8/8/2019 3:34:19 PM	3150	0.1	1	1.1	0.00	51
CO Line 2							
	8/8/2019 3:42:49 PM	3660	0.1	0	0.1	0.00	206
	8/8/2019 3:43:19 PM	3690	0.1	1	1.1	0.00	206
	8/8/2019 3:43:49 PM	3720	0.1	0	0.1	0.00	206
	8/8/2019 3:44:19 PM	3750	0.1	1	1.1	0.00	206
	8/8/2019 3:44:49 PM	3780	0.1	1	1.1	0.00	206
	8/8/2019 3:45:19 PM	3810	0.1	1	1.1	0.00	206
	8/8/2019 3:45:49 PM	3840	0.1	0	0.1	0.00	206
	8/8/2019 3:46:19 PM	3870	0.1	0	0.1	0.00	206
	8/8/2019 3:46:49 PM	3900	0.1	1	1.1	0.00	206
	8/8/2019 3:47:19 PM	3930	0.1	0	0.1	0.00	206

Company:	Equitrans Midstream	Span:	15.04
Station:	McIntosh	Technician:	C.Ent
Unit:	4	Analyzer:	Testo 350
Date:	8/27/2019	Anal. Ser. No:	60427447
		Method:	3A
Cal Gas S/N	<b>CC515961</b>		

**O2**

	Gas Conc. ppm	Anal. Cal Response	Cal. Error % Span	System Cal Response	System Cal. bias(%span)	Drift%
Pre Run						
Zero	<b>0.00</b>	0.00	<b>0.0</b>	0.0	0.0	-
High	15.04	14.85	<b>-1.3</b>	14.8	0.0	-
Post Run 1						
Zero	0.00	0.00	N/A	0.00	0.0	0.00
Upscale	15.04	14.83	N/A	14.83	<b>0.0</b>	<b>-0.15</b>

Calibration error = ( (Anal. Cal Response - Gas Conc.)/Instrument span) x 100

Line Bias = ((System Cal Response - Anal. Cal Response)/Instrument span) x 100

Drift = ((Current System Observed Response-Initial System Cal Response)/Span) x 100

## Corrections

Run #	Average	Co	Cm	Cma	Corrected
1	10.88	0.00	14.84	15.04	<b>11.03</b>

Corrected = (Run Ave.-Co) x (Cma/Cm-Co)

Run Ave. = Average of the test run

Co = Ave. of initial & current zero system response

Cm = Ave. of initial & current upscale system response

Cma = Actual upscale gas concentration



Company:	Equitrans Midstream	Span:	51.15
Station:	McIntosh	Technician:	C.Ent
Unit:	4	Analyzer:	Testo 350
Date:	8/27/2019	Anal. Ser. No:	60427447
		Method:	7E
Cal Gas S/N	<b>CC515961</b>		

**NO2**

	Gas Conc. ppm	Anal. Cal Response	Cal. Error % Span	System Cal Response	System Cal. bias(%span)	Drift%
Pre Run						
Zero	0.00	0.0	0.0	0.0	0.0	-
High	51.15	50.2	-2.0	50.2	0.0	-
Post Run 1						
Zero	0.0	0.0	N/A	0.0	0.0	0.0
Upscale	51.15	49.8	N/A	49.8	0.0	-0.7

Calibration error = (Anal. Cal Response - Gas Conc.)/Instrument span) x 100

Line Bias = ((System Cal Response - Anal. Cal Response)/Instrument span) x 100

Drift = ((Current System Observed Response-Initial System Cal Response)/Span) x 100

## Corrections

Run #	Average	Co	Cm	Cma	Corrected
1	9.1	0.0	50.0	51.2	<b>9.30</b>

Corrected = (Run Ave.-Co) x (Cma/Cm-Co)

Run Ave. = Average of the test run

Co = Ave. of initial & current zero system response

Cm = Ave. of initial & current upscale system response

Cma = Actual upscale gas concentration

Company:	Equitrans Midstream	Span:	50.76
Station:	McIntosh	Technician:	C.Ent
Unit:	4	Analyzer:	Testo 350
Date:	8/27/2019	Anal. Ser. No:	60427447
		Method:	7E

Cal Gas S/N	<b>CC1205</b>	<b>NO interference Response %</b>	<b>0.1</b>
		<b>Must be ≤ 5%</b>	

NO

	Gas Conc. ppm	Anal. Cal Response	Cal. Error % Span	System Cal Response	System Cal. bias(%span)	Drift%
Pre Run						
Zero	0.00	0.00	0.0	0.0	0.0	-
High	50.76	50.00	-1.5	50.0	0.0	-
Post Run 1						
Zero	0.00	0.00	N/A	0.00	0.0	0.00
Upscale	50.76	49.00	N/A	49.00	0.0	-1.97

Calibration error =  $(\text{Anal. Cal Response} - \text{Gas Conc.}) / \text{Instrument span} \times 100$

Line Bias =  $(\text{System Cal Response} - \text{Anal. Cal Response}) / \text{Instrument span} \times 100$

Drift =  $(\text{Current System Observed Response} - \text{Initial System Cal Response}) / \text{Span} \times 100$

## Corrections

Run #	Average	Co	Cm	Cma	Corrected
1	21.6	0.0	49.5	50.8	<b>22.19</b>

Corrected =  $(\text{Run Ave.} - \text{Co}) \times (\text{Cma} / \text{Cm} - \text{Co})$

Run Ave. = Average of the test run

Co = Ave. of initial & current zero system response

Cm = Ave. of initial & current upscale system response

Cma = Actual upscale gas concentration

Company:	Equitrans Midstream	Span:	51.19
Station:	McIntosh	Technician:	C.Ent
Unit:	4	Analyzer:	Testo 350
Date:	8/27/2019	Anal. Ser. No:	60427447
		Method:	10
Cal Gas S/N	<b>CC186134</b>		

<b>CO interference Response %</b>	<b>0.0</b>
<b>Must be ≤ 5%</b>	

**CO**

	Gas Conc. ppm	Anal. Cal Response	Cal. Error % Span	System Cal Response	System Cal. bias(%span)	Drift%
Pre Run						
Zero	0.00	1.00	2.0	1.0	0.0	-
High	51.19	52.00	1.6	52.00	0.0	-
Post Run 1						
Zero	0.00	1.00	N/A	1.00	0.0	0.00
Upscale	51.19	51.25	N/A	51.25	0.0	-1.47

Calibration error = (Anal. Cal Response - Gas Conc.)/Instrument span) x 100

Line Bias = ((System Cal Response - Anal. Cal Response)/Instrument span) x 100

Drift = ((Current System Observed Response-Initial System Cal Response)/Span) x 100

**Corrections**

Run #	Average	Co	Cm	Cma	Corrected
1	5.8	1.00	51.63	51.19	<b>4.87</b>

Corrected = (Run Ave.-Co) x (Cma/Cm-Co)

Run Ave. = Average of the test run

Co = Ave. of initial & current zero system response

Cm = Ave. of initial & current upscale system response

Cma = Actual upscale gas concentration

## Section IV.

### Site Data

**Data input**

Station:	McIntosh	Manfct:	Caterpillar	<a href="#">Enter Field data in all cells that are blue</a>							
Engine #:	4	Model:	3616 A4								
HP:	5350	Rated Speed:	1000								
Run #:	1	Date:	8/27/2019						Analyzer Testo 350		
		Permit hours:	8760						S/N 60427447		
Data Collected by:	C.Ent	Cylinder ID	Calibration Data	Cal.Bottle	Pre-test	Post-test	Pre-test zero	Post-test zero	R <sub>NO-NO2</sub>	R <sub>CO-NO2</sub>	R <sub>CO-NO</sub>
2 or 4 Stroke	4	CC515961	O2 span (%)	15.04	14.85	14.83	0.00	0.00			
Bore (in)	11.8	CC1205	NO span (ppm)	50.76	50.00	49.00	0.00	0.00	0.10		
Stroke (in)	11.8	CC515961	NO2 span (ppm)	51.15	50.15	49.80	0.00	0.00			
# Cylinders	16	CC186134	CO span (ppm)	51.2	52.0	51.3	1.00	1.00		0.00	0.00
Exh Port Close BTDC	291										Post test interference verf.
Rod Length (in)	23.622										

Engine Data		Date	Time	SAMPLE #	RPM	HP	Ign. Timing	AMT degF	AMP "Hg	FFR (mscfh)	FFR (scfm)	% Load
		8/27/2019	11:08 AM	1	1000	4975.5	18.1	108	35	36.700	611.667	93.0%
		8/27/2019	11:23 AM	2	1000	4815.0	17.9	108	33	35.800	596.667	90.0%
		8/27/2019	11:38 AM	3	998	4761.5	17.9	108	33	35.600	593.333	89.0%
		8/27/2019	11:53 AM	4	1000	4815.0	17.9	109	33	35.600	593.333	90.0%
		8/27/2019	12:08 PM	5	1000	4761.5	17.9	109	33	35.600	593.333	89.0%
				<b>AVERAGES</b>	<b>999.6</b>	<b>4825.7</b>	<b>17.9</b>	<b>108.4</b>	<b>33.3</b>	<b>35.860</b>	<b>597.67</b>	<b>90.2%</b>

Run Averages	NO2 ppm	NO ppm	NOx ppm	O2 %	CO ppm	Pump L/m
	9.1	22	31	10.88	6	0.97
Baramatric pressure	29.95 in		Amb Temp	68.00 F		
Humidity	0.94 %					

**GAS COMPOSITION**

Constituent	Mole%	Upper Dry Heat Value	1042.0	btu/dscf/14.73psia
NITROGEN	0.2879	Low Dry Heat Value	948.2	btu/dscf
CARBON DIOX.	0.4846	Specific Gravity	0.5842	
METHANE	94.6868	Oxygen	0.0001	
ETHANE	4.154	Carbon Monoxide	0.0000	
PROPANE	0.3038	Hydrogen	0.0000	
I-BUTANE	0.0191	Heptane	0.0000	
N-BUTANE	0.0386	H2S	0.0000	
I-PENTANE	0	SO2	0.0000	
N-PENTANE	0.0251	He	0.0000	
HEXANE +	0	F Factor calc	8633.8	
	100.0000	F Factor Method 19	8710.0	

## Section V.

### Run Data

Date / time	sec Runtime	ppm NO2	ppm NO	ppm NOx	% O2	ppm CO
8/27/2019 11:08:08 AM	0	7.4	20	27.4	10.94	6
8/27/2019 11:08:38 AM	30	9.2	25	34.2	10.77	6
8/27/2019 11:09:08 AM	60	9.1	24	33.1	10.83	6
8/27/2019 11:09:38 AM	90	9.2	24	33.2	10.81	6
8/27/2019 11:10:08 AM	120	9.0	23	32.0	10.88	6
8/27/2019 11:10:38 AM	150	8.7	23	31.7	10.86	6
8/27/2019 11:11:08 AM	180	8.5	23	31.5	10.99	6
8/27/2019 11:11:38 AM	210	9.0	24	33.0	10.85	6
8/27/2019 11:12:08 AM	240	9.0	24	33.0	10.86	6
8/27/2019 11:12:38 AM	270	8.9	23	31.9	10.98	6
8/27/2019 11:13:08 AM	300	9.1	23	32.1	10.87	6
8/27/2019 11:13:38 AM	330	9.0	23	32.0	10.85	6
8/27/2019 11:14:08 AM	360	9.0	22	31.0	10.77	6
8/27/2019 11:14:38 AM	390	9.3	23	32.3	10.84	6
8/27/2019 11:15:08 AM	420	9.3	23	32.3	11.09	6
8/27/2019 11:15:38 AM	450	9.3	24	33.3	10.77	6
8/27/2019 11:16:08 AM	480	9.3	23	32.3	10.85	6
8/27/2019 11:16:38 AM	510	9.5	24	33.5	10.84	6
8/27/2019 11:17:08 AM	540	9.7	24	33.7	10.82	6
8/27/2019 11:17:38 AM	570	9.7	24	33.7	10.83	6
8/27/2019 11:18:08 AM	600	8.6	21	29.6	10.87	6
8/27/2019 11:18:38 AM	630	8.1	19	27.1	11.01	6
8/27/2019 11:19:08 AM	660	10.1	26	36.1	10.85	6
8/27/2019 11:19:38 AM	690	9.7	24	33.7	10.83	6
8/27/2019 11:20:08 AM	720	10.0	25	35.0	10.82	6
8/27/2019 11:20:38 AM	750	9.4	23	32.4	10.82	6
8/27/2019 11:21:08 AM	780	9.5	23	32.5	10.85	6
8/27/2019 11:21:38 AM	810	9.3	23	32.3	10.87	5
8/27/2019 11:22:08 AM	840	9.3	23	32.3	10.85	5
8/27/2019 11:22:38 AM	870	9.3	22	31.3	10.86	5
8/27/2019 11:23:08 AM	900	9.4	22	31.4	10.88	6
8/27/2019 11:23:38 AM	930	9.1	22	31.1	10.89	6
8/27/2019 11:24:08 AM	960	9.0	21	30.0	10.91	5
8/27/2019 11:24:38 AM	990	8.9	21	29.9	10.88	6
8/27/2019 11:25:08 AM	1020	8.9	21	29.9	10.90	6
8/27/2019 11:25:38 AM	1050	9.3	22	31.3	10.85	6
8/27/2019 11:26:08 AM	1080	9.3	22	31.3	10.87	6
8/27/2019 11:26:38 AM	1110	9.3	22	31.3	10.86	6
8/27/2019 11:27:08 AM	1140	9.5	22	31.5	10.70	5
8/27/2019 11:27:38 AM	1170	9.6	23	32.6	10.83	5
8/27/2019 11:28:08 AM	1200	9.8	23	32.8	10.82	6
8/27/2019 11:28:38 AM	1230	9.5	23	32.5	10.79	6
8/27/2019 11:29:08 AM	1260	9.3	22	31.3	10.88	6
8/27/2019 11:29:38 AM	1290	9.3	22	31.3	10.83	6
8/27/2019 11:30:08 AM	1320	9.6	23	32.6	10.78	6
8/27/2019 11:30:38 AM	1350	9.3	22	31.3	10.90	6
8/27/2019 11:31:08 AM	1380	9.5	22	31.5	10.86	6
8/27/2019 11:31:38 AM	1410	9.5	23	32.5	10.97	6
8/27/2019 11:32:08 AM	1440	9.9	24	33.9	10.81	5
8/27/2019 11:32:38 AM	1470	9.9	24	33.9	10.87	5

8/27/2019 11:33:08 AM	1500	9.6	23	32.6	10.79	6
8/27/2019 11:33:38 AM	1530	9.0	21	30.0	10.88	6
8/27/2019 11:34:08 AM	1560	8.9	21	29.9	10.88	6
8/27/2019 11:34:38 AM	1590	8.3	20	28.3	10.94	6
8/27/2019 11:35:08 AM	1620	9.1	21	30.1	10.83	6
8/27/2019 11:35:38 AM	1650	9.7	25	34.7	10.79	5
8/27/2019 11:36:08 AM	1680	7.9	19	26.9	10.96	6
8/27/2019 11:36:38 AM	1710	8.8	24	32.8	10.90	6
8/27/2019 11:37:08 AM	1740	9.3	24	33.3	10.84	6
8/27/2019 11:37:38 AM	1770	8.8	24	32.8	10.84	6
8/27/2019 11:38:08 AM	1800	7.9	18	25.9	10.94	6
8/27/2019 11:38:38 AM	1830	8.1	18	26.1	10.91	6
8/27/2019 11:39:08 AM	1860	8.0	18	26.0	10.97	6
8/27/2019 11:39:38 AM	1890	8.3	20	28.3	10.97	6
8/27/2019 11:40:08 AM	1920	8.4	19	27.4	10.98	6
8/27/2019 11:40:38 AM	1950	9.8	25	34.8	10.75	5
8/27/2019 11:41:08 AM	1980	9.8	25	34.8	10.74	5
8/27/2019 11:41:38 AM	2010	10.0	26	36.0	10.91	6
8/27/2019 11:42:08 AM	2040	9.0	24	33.0	10.81	6
8/27/2019 11:42:38 AM	2070	7.9	18	25.9	10.98	6
8/27/2019 11:43:08 AM	2100	8.0	18	26.0	11.09	6
8/27/2019 11:43:38 AM	2130	9.5	23	32.5	10.80	6
8/27/2019 11:44:08 AM	2160	9.8	24	33.8	10.85	6
8/27/2019 11:44:38 AM	2190	8.2	19	27.2	10.92	6
8/27/2019 11:45:08 AM	2220	7.9	17	24.9	11.03	6
8/27/2019 11:45:38 AM	2250	9.3	24	33.3	10.83	6
8/27/2019 11:46:08 AM	2280	8.4	18	26.4	10.95	6
8/27/2019 11:46:38 AM	2310	9.0	21	30.0	10.87	6
8/27/2019 11:47:08 AM	2340	8.7	20	28.7	10.90	6
8/27/2019 11:47:38 AM	2370	8.8	20	28.8	10.88	6
8/27/2019 11:48:08 AM	2400	8.7	20	28.7	10.89	5
8/27/2019 11:48:38 AM	2430	8.8	20	28.8	10.90	6
8/27/2019 11:49:08 AM	2460	8.7	19	27.7	10.91	6
8/27/2019 11:49:38 AM	2490	8.7	20	28.7	10.91	6
8/27/2019 11:50:08 AM	2520	8.7	19	27.7	10.91	6
8/27/2019 11:50:38 AM	2550	9.0	20	29.0	10.88	6
8/27/2019 11:51:08 AM	2580	8.9	20	28.9	10.89	6
8/27/2019 11:51:38 AM	2610	9.1	21	30.1	10.89	5
8/27/2019 11:52:08 AM	2640	8.9	20	28.9	10.90	6
8/27/2019 11:52:38 AM	2670	8.8	20	28.8	10.88	5
8/27/2019 11:53:08 AM	2700	8.4	19	27.4	10.92	6
8/27/2019 11:53:38 AM	2730	8.6	17	25.6	10.92	6
8/27/2019 11:54:08 AM	2760	9.8	23	32.8	10.82	6
8/27/2019 11:54:38 AM	2790	8.1	18	26.1	10.97	6
8/27/2019 11:55:08 AM	2820	8.8	21	29.8	10.96	6
8/27/2019 11:55:38 AM	2850	8.6	19	27.6	10.98	6
8/27/2019 11:56:08 AM	2880	8.4	18	26.4	10.97	6
8/27/2019 11:56:38 AM	2910	9.4	20	29.4	10.83	6
8/27/2019 11:57:08 AM	2940	10.1	24	34.1	10.79	5
8/27/2019 11:57:38 AM	2970	8.6	21	29.6	10.98	6
8/27/2019 11:58:08 AM	3000	9.6	22	31.6	10.88	6



8/27/2019 11:58:38 AM	3030	9.9	24	33.9	10.79	5
8/27/2019 11:59:08 AM	3060	9.9	23	32.9	10.85	5
8/27/2019 11:59:38 AM	3090	9.5	23	32.5	10.82	6
8/27/2019 12:00:08 PM	3120	9.3	21	30.3	10.89	6
8/27/2019 12:00:38 PM	3150	9.2	21	30.2	10.90	6
8/27/2019 12:01:08 PM	3180	9.1	21	30.1	10.90	6
8/27/2019 12:01:38 PM	3210	8.8	20	28.8	10.90	6
8/27/2019 12:02:08 PM	3240	8.9	20	28.9	10.92	6
8/27/2019 12:02:38 PM	3270	8.8	19	27.8	10.91	6
8/27/2019 12:03:08 PM	3300	9.0	20	29.0	10.90	6
8/27/2019 12:03:38 PM	3330	9.0	20	29.0	10.89	5
8/27/2019 12:04:08 PM	3360	9.0	20	29.0	10.89	6
8/27/2019 12:04:38 PM	3390	9.1	20	29.1	10.90	5
8/27/2019 12:05:08 PM	3420	9.0	20	29.0	10.89	6
8/27/2019 12:05:38 PM	3450	9.1	20	29.1	10.88	5
8/27/2019 12:06:08 PM	3480	8.8	20	28.8	10.90	6
8/27/2019 12:06:38 PM	3510	8.2	19	27.2	11.00	6
8/27/2019 12:07:08 PM	3540	8.4	18	26.4	10.94	6
8/27/2019 12:07:38 PM	3570	9.9	23	32.9	10.84	5
8/27/2019 12:08:08 PM	3600	8.7	18	26.7	10.94	6
8/27/2019 12:08:38 PM	3630	9.7	23	32.7	10.80	5
8/27/2019 12:09:08 PM	3660	10.1	24	34.1	10.83	6
8/27/2019 12:09:38 PM	3690	10.2	24	34.2	10.85	6
8/27/2019 12:10:08 PM	3720	9.8	24	33.8	10.76	5
8/27/2019 12:10:38 PM	3750	9.6	22	31.6	10.86	6
8/27/2019 12:11:08 PM	3780	10.0	23	33.0	10.82	6
8/27/2019 12:11:38 PM	3810	8.4	21	29.4	11.08	6
		<b>9.1</b>	<b>22</b>	<b>31</b>	<b>10.88</b>	<b>6</b>

## Section VI.

### Fuel Gas Analysis



EQT Midstream  
 555 Dishong Mountain Road  
 Johnstown Pa. 15906

Run Date 8/28/2019  
 Run Time 3:12 PM

## Fractional Analysis

Station	McIntosh	Sample Date	8/27/19
		Sample Time	1:20 PM
Meter	McIntosh	Sample Collected	B.C.Ent

Component	MOL %	GPM	Analytical Results At Base Conditions (Real)	
Methane	94.6868		BTU/SCF (Dry)	1041.9999
Ethane	4.1540	1.11	BTU/SCF (Saturated)	1023.8322
Propane	0.3038	0.08	PSIA	14.73
I-Butane	0.0191	0.01	Temperature (F)	60.00
N-Butane	0.0386	0.01		
I-Pentane	0.0000	0.00		
N-Pentane	0.0251	0.01		
Nitrogen	0.2879			
CO2	0.4846		Calculated Specific Gravities	
Oxygen	0.0001		Real Gravity	0.5842
Hexanes+	0.0000	0.00	Molecular Weight	16.9188
<b>Total:</b>	<b>100.0000</b>			

## Section VII.

### Calibration Gas Certificates



an Air Liquide company

**Airgas Specialty Gases**Airgas USA, LLC  
12722 S. Wentworth Ave.  
Chicago, IL 60628  
Airgas.com

# CERTIFICATE OF ANALYSIS

## Grade of Product: EPA Protocol

Part Number:	E02NI99E33A0284	Reference Number:	54-124451484-1
Cylinder Number:	SG9101757ALD1	Cylinder Volume:	31.6 CF
Laboratory:	124 - Chicago (SAP) - IL	Cylinder Pressure:	2217 PSIG
PGVP Number:	B12014	Valve Outlet:	660
Gas Code:	NO,NOX,BALN	Certification Date:	Sep 19, 2014

**Expiration Date: Sep 19, 2022**

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

**ANALYTICAL RESULTS**

Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty	Assay Dates
NOX	200.0 PPM	199.9 PPM	G1	+/- 0.8% NIST Traceable	09/12/2014, 09/19/2014
NITRIC OXIDE	200.0 PPM	199.9 PPM	G1	+/- 0.8% NIST Traceable	09/12/2014, 09/19/2014
NITROGEN	Balance			-	

**CALIBRATION STANDARDS**

Type	Lot ID	Cylinder No	Concentration	Uncertainty	Expiration Date
NTRM	12061936	CC367646	250.8 PPM NITRIC OXIDE/NITROGEN	+/- 0.5%	May 04, 2018
PRM	12312	680179	10.01 PPM NITROGEN DIOXIDE/NITROGEN	+/- 2.0%	Oct 15, 2014
GMIS	124206889102	CC320508	4.979 PPM NITROGEN DIOXIDE/NITROGEN	+/- 2.0%	May 04, 2015

The SRM, PRM or RGM noted above is only in reference to the GMIS used in the assay and not part of the analysis.

**ANALYTICAL EQUIPMENT**

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration

Triad Data Available Upon Request



Signature on file

Approved for Release



an Air Liquide company

Airgas Specialty Gases

Airgas USA, LLC  
12722 S. Wentworth Ave.  
Chicago, IL 60628  
Airgas.com

# CERTIFICATE OF ANALYSIS

## Grade of Product: EPA Protocol

Part Number:	E02NI99E33A0284	Reference Number:	54-124445125-1
Cylinder Number:	FF48921	Cylinder Volume:	31.6 CF
Laboratory:	124 - Chicago (SAP) - IL	Cylinder Pressure:	2217 PSIG
PGVP Number:	B12014	Valve Outlet:	660
Gas Code:	NO,NOX,BALN	Certification Date:	Aug 01, 2014

**Expiration Date: Aug 01, 2022**

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

### ANALYTICAL RESULTS

Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty	Assay Dates
NOX	200.0 PPM	197.6 PPM	G1	+/- 0.7% NIST Traceable	07/25/2014, 08/01/2014
NITRIC OXIDE	200.0 PPM	197.6 PPM	G1	+/- 0.7% NIST Traceable	07/25/2014, 08/01/2014
NITROGEN	Balance			-	

### CALIBRATION STANDARDS

Type	Lot ID	Cylinder No	Concentration	Uncertainty	Expiration Date
NTRM	12061936	CC367646	250.8 PPM NITRIC OXIDE/NITROGEN	+/- 0.5%	May 04, 2018
PRM	12312	680179	10.01 PPM NITROGEN DIOXIDE/NITROGEN	+/- 2.0%	Oct 15, 2014
GMIS	124206889102	CC320508	4.979 PPM NITROGEN DIOXIDE/NITROGEN	+/- 2.0%	May 04, 2015

The SRM, PRM or RGM noted above is only in reference to the GMIS used in the assay and not part of the analysis.

### ANALYTICAL EQUIPMENT

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration

Triad Data Available Upon Request



Signature on file

Approved for Release



an Air Liquide company

Airgas Great Lakes region

Airgas USA, LLC  
1290 Combermere Dr.  
Troy, MI 48083  
Airgas.com

# CERTIFICATE OF ANALYSIS

## Grade of Product: EPA Protocol

Customer: EQT MIDSTREAM -  
LANE JOBE

Part Number: E02NI99E33A0919 Reference Number: 32-401426337-1

Cylinder Number: FF14201 Cylinder Volume: 31.6 CF

Laboratory: 112 - Troy-32 (SAP) - MI Cylinder Pressure: 2217 PSIG

PGVP Number: B62019 Valve Outlet: 660

Gas Code: NO,NOX,BALN Certification Date: Mar 07, 2019

**Expiration Date: Mar 07, 2027**

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

**ANALYTICAL RESULTS**

Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty	Assay Dates
NOX	500.0 PPM	501.8 PPM	G1	+/- 0.7% NIST Traceable	02/28/2019, 03/07/2019
NITRIC OXIDE	500.0 PPM	501.0 PPM	G1	+/- 0.7% NIST Traceable	02/28/2019, 03/07/2019
NITROGEN	Balance			-	

**CALIBRATION STANDARDS**

Type	Lot ID	Cylinder No	Concentration	Uncertainty	Expiration Date
NTRM	15010102	KAL003004	494.6 PPM NITRIC OXIDE/NITROGEN	+/-0.5%	Sep 01, 2021

**ANALYTICAL EQUIPMENT**

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
E/N 54 Nicolet 6700 NO	FTIR	Feb 12, 2019
E/N 54 Nicolet 6700 NO2	FTIR	Feb 12, 2019

Triad Data Available Upon Request



Signature on file

Approved for Release

Page 1 of 32-401426337-1

**CERTIFICATE OF ANALYSIS****Grade of Product: EPA Protocol**

Customer: EQT MIDSTREAM  
 Part Number: E02NI99E33A0551  
 Cylinder Number: EA0011073  
 Laboratory: 112 - Royal Oak-32 (SAP) - MI  
 PGVP Number: B62015  
 Gas Code: CO,BALN

Reference Number: 32-400636061-1  
 Cylinder Volume: 28.5 CF  
 Cylinder Pressure: 1984 PSIG  
 Valve Outlet: 350  
 Certification Date: Dec 07, 2015

**Expiration Date: Dec 07, 2023**

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

<b>ANALYTICAL RESULTS</b>					
Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty	Assay Dates
CARBON MONOXIDE	200.0 PPM	203.3 PPM	G1	+/- 0.5% NIST Traceable	12/07/2015
NITROGEN	Balance			-	

<b>CALIBRATION STANDARDS</b>					
Type	Lot ID	Cylinder No	Concentration	Uncertainty	Expiration Date
NTRM	14060316	CC432081	252.5 PPM CARBON MONOXIDE/NITROGEN	+/-0.3%	Feb 21, 2020

<b>ANALYTICAL EQUIPMENT</b>		
Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
E/N 54 Nicolet 6700 CO	FTIR	Nov 18, 2015

Triad Data Available Upon Request

PERMANENT NOTES:Produced at: 2009 Bellaire Ave, Royal Oak, MI 48067



Signature on file

Approved for Release





an Air Liquide company

Airgas Specialty Gases

Airgas USA, LLC  
630 United Drive  
Durham, NC 27713  
Airgas.com

# CERTIFICATE OF ANALYSIS

## Grade of Product: EPA Protocol

Part Number:	E03NI84E33W0000	Reference Number:	122-401034280-1
Cylinder Number:	D496834	Cylinder Volume:	31.9 Cubic Feet
Laboratory:	124 - Durham (SAP) - NC	Cylinder Pressure:	2216 PSIG
PGVP Number:	B22017	Valve Outlet:	660
Gas Code:	NO2,O2,BALN	Certification Date:	Oct 20, 2017

**Expiration Date: Oct 20, 2020**

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

**ANALYTICAL RESULTS**

Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty	Assay Dates
NITROGEN DIOXIDE	100.0 PPM	95.73 PPM	G1	+/- 1.2% NIST Traceable	10/13/2017, 10/20/2017
OXYGEN	15.00 %	15.06 %	G1	+/- 0.3% NIST Traceable	10/17/2017
NITROGEN	Balance			-	

**CALIBRATION STANDARDS**

Type	Lot ID	Cylinder No	Concentration	Uncertainty	Expiration Date
GMIS	124542144102	CC507257	197.1 PPM NITROGEN DIOXIDE/NITROGEN	+/- 1.1%	Mar 13, 2020
PRM	12370	FF27640	200.2 PPM NITROGEN DIOXIDE/NITROGEN	+/- 1.0%	Oct 13, 2017

The SRM, PRM or RGM noted above is only in reference to the GMIS used in the assay and not part of the analysis.

**ANALYTICAL EQUIPMENT**

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
THERMO 48I CO 1308857346	Nondispersive Infrared (NDIR)	Oct 04, 2017
Horiba MPA510 O2 41499150042	Paramagnetic	Oct 05, 2017

Triad Data Available Upon Request



Signature on file

Approved for Release



an Air Liquide company

Airgas Specialty Gases

Airgas USA, LLC  
12722 S. Wentworth Ave.  
Chicago, IL 60628  
Airgas.com

# CERTIFICATE OF ANALYSIS

## Grade of Product: EPA Protocol

Part Number:	E03NI84E15W0000	Reference Number:	54-401406062-1
Cylinder Number:	CC515961	Cylinder Volume:	145.7 CF
Laboratory:	124 - Chicago (SAP) - IL	Cylinder Pressure:	2015 PSIG
PGVP Number:	B12019	Valve Outlet:	660
Gas Code:	NO2,O2,BALN	Certification Date:	Feb 11, 2019

**Expiration Date: Feb 11, 2022**

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

### ANALYTICAL RESULTS

Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty	Assay Dates
NITROGEN DIOXIDE	50.00 PPM	51.15 PPM	G1	+/- 2.0% NIST Traceable	01/28/2019, 02/11/2019
OXYGEN	15.00 %	15.04 %	G1	+/- 0.8% NIST Traceable	01/29/2019
NITROGEN	Balance			-	

### CALIBRATION STANDARDS

Type	Lot ID	Cylinder No	Concentration	Uncertainty	Expiration Date
GMIS	7282017108	CC511467	60.62 PPM NITROGEN DIOXIDE/NITROGEN	+/- 1.6%	Sep 03, 2021
PRM	12378	D562913	100.1 PPM NITROGEN DIOXIDE/AIR	+/- 1.0%	Sep 04, 2018
NTRM	15010409	K013750	22.454 % OXYGEN/NITROGEN	+/- 0.2%	Aug 05, 2021

The SRM, PRM or RGM noted above is only in reference to the GMIS used in the assay and not part of the analysis.

### ANALYTICAL EQUIPMENT

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
MKS FTIR NO2 017707558	FTIR	Feb 05, 2019
O2-1 HORIBA MPA-510 3VUYL9NR	Paramagnetic	Jan 21, 2019

Triad Data Available Upon Request



Signature on file

Approved for Release



an Air Liquide company

Airgas Specialty Gases

Airgas USA, LLC  
12722 S. Wentworth Ave.  
Chicago, IL 60628  
Airgas.com

# CERTIFICATE OF ANALYSIS

## Grade of Product: EPA Protocol

Part Number:	E03NI84E15W0007	Reference Number:	54-401029259-1
Cylinder Number:	CC507258	Cylinder Volume:	145.7 Cubic Feet
Laboratory:	124 - Chicago (SAP) - IL	Cylinder Pressure:	2015 PSIG
PGVP Number:	B12017	Valve Outlet:	660
Gas Code:	NO2,O2,BALN	Certification Date:	Oct 30, 2017

**Expiration Date: Oct 30, 2020**

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

### ANALYTICAL RESULTS

Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty	Assay Dates
NITROGEN DIOXIDE	100.0 PPM	101.3 PPM	G1	+/- 2% NIST Traceable	10/23/2017, 10/30/2017
OXYGEN	15.00 %	15.05 %	G1	+/- 0.5% NIST Traceable	10/23/2017
NITROGEN	Balance			-	

### CALIBRATION STANDARDS

Type	Lot ID	Cylinder No	Concentration	Uncertainty	Expiration Date
GMIS	0120201702	CC503906	101.3 PPM NITROGEN DIOXIDE/NITROGEN	+/- 1.1%	Jan 20, 2020
PRM	MS7323	MS7323	100.1 PPM NITROGEN DIOXIDE/NITROGEN	+/- 1.0%	May 19, 2016
NTRM	09061418	CC273593	22.53 % OXYGEN/NITROGEN	+/- 0.4%	Mar 08, 2019

The SRM, PRM or RGM noted above is only in reference to the GMIS used in the assay and not part of the analysis.

### ANALYTICAL EQUIPMENT

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
MKS Multigas 017707558	FTIR	Oct 04, 2017
O2-1 HORIBA MPA-510 3VUYL9NR	Paramagnetic	Sep 29, 2017

Triad Data Available Upon Request



Signature on file

Approved for Release



an Air Liquide company

**Airgas Specialty Gases**Airgas USA, LLC  
12722 S. Wentworth Ave.  
Chicago, IL 60628  
Airgas.com**CERTIFICATE OF ANALYSIS****Grade of Product: EPA Protocol**

Part Number:	E02NI99E15A0302	Reference Number:	54-124299381-1
Cylinder Number:	CC186134	Cylinder Volume:	144.3 Cubic Feet
Laboratory:	124 - Chicago (SAP) - IL	Cylinder Pressure:	2015 PSIG
PGVP Number:	B12012	Valve Outlet:	350
Gas Code:	CO,BALN	Certification Date:	Jan 24, 2012

**Expiration Date: Jan 24, 2020**

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

**ANALYTICAL RESULTS**

Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty	Assay Dates
CARBON MONOXIDE	50.00 PPM	51.19 PPM	G1	+/- 1% NIST Traceable	01/13/2012, 01/24/2012
NITROGEN	Balance			-	

**CALIBRATION STANDARDS**

Type	Lot ID	Cylinder No	Concentration	Uncertainty	Expiration Date
NTRM/CO	05120309	CC179906	49.33 PPM CARBON MONOXIDE/NITROGEN	+/- 0.7%	Feb 02, 2013

**ANALYTICAL EQUIPMENT**

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
CO-1 SIEMENS ULTRAMAT 6E N1J5700	NDIR	Dec 26, 2011

Triad Data Available Upon Request



Signature on file

Approved for Release



an Air Liquide company

Airgas Great Lakes region

Airgas USA, LLC  
1290 Combermere Dr.  
Troy, MI 48083  
Airgas.com

# CERTIFICATE OF ANALYSIS

## Grade of Product: EPA Protocol

Customer:	EQT MIDSTREAM	Reference Number:	32-401041391-1
Part Number:	E02NI96E15A3095	Cylinder Volume:	144.7 CF
Cylinder Number:	CC136761	Cylinder Pressure:	2015 PSIG
Laboratory:	112 - Troy-32 (SAP) - MI	Valve Outlet:	580
PGVP Number:	B62017	Certification Date:	Nov 14, 2017
Gas Code:	O2,BALN		

**Expiration Date: Nov 14, 2025**

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

ANALYTICAL RESULTS					
Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty	Assay Dates
OXYGEN	4.000 %	4.007 %	G1	+/- 0.5% NIST Traceable	11/14/2017
NITROGEN	Balance			-	

CALIBRATION STANDARDS					
Type	Lot ID	Cylinder No	Concentration	Uncertainty	Expiration Date
NTRM	140606-02	CC431989	4.794 % OXYGEN/NITROGEN	+/-0.4%	Nov 29, 2019

ANALYTICAL EQUIPMENT		
Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
O2 FS, SIEMENS OXYMAT 6 E/N 182	Paramagnetic	Oct 19, 2017

Triad Data Available Upon Request



Signature on file

Approved for Release

**CERTIFICATE OF ANALYSIS****Grade of Product: EPA Protocol**

Customer: EQT MIDSTREAM  
 Part Number: E02NI99E15A0404  
 Cylinder Number: CC6889  
 Laboratory: 112 - Royal Oak-32 (SAP) - MI  
 PGVP Number: B62016  
 Gas Code: CO,BALN

Reference Number: 32-400752406-1  
 Cylinder Volume: 144.3 CF  
 Cylinder Pressure: 2015 PSIG  
 Valve Outlet: 350  
 Certification Date: Aug 04, 2016

**Expiration Date: Aug 04, 2024**

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

<b>ANALYTICAL RESULTS</b>					
Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty	Assay Dates
CARBON MONOXIDE	200.0 PPM	200.7 PPM	G1	+/- 0.5% NIST Traceable	08/04/2016
NITROGEN	Balance			-	

<b>CALIBRATION STANDARDS</b>					
Type	Lot ID	Cylinder No	Concentration	Uncertainty	Expiration Date
NTRM	14060316	CC432081	252.5 PPM CARBON MONOXIDE/NITROGEN	+/-0.3%	Feb 21, 2020

<b>ANALYTICAL EQUIPMENT</b>		
Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
E/N 54 Nicolet 6700 CO	FTIR	Aug 01, 2016

Triad Data Available Upon Request

PERMANENT NOTES:Produced at: 2009 Bellaire Ave, Royal Oak, MI 48067



Signature on file

Approved for Release



an Air Liquide company

**Airgas Specialty Gases**Airgas USA, LLC  
12722 S. Wentworth Ave.  
Chicago, IL 60628  
Airgas.com**CERTIFICATE OF ANALYSIS****Grade of Product: EPA Protocol**

Part Number:	E02NI99E15A0147	Reference Number:	54-401409450-1
Cylinder Number:	CC1205	Cylinder Volume:	144.3 CF
Laboratory:	124 - Chicago (SAP) - IL	Cylinder Pressure:	2015 PSIG
PGVP Number:	B12019	Valve Outlet:	660
Gas Code:	NO,NOX,BALN	Certification Date:	Feb 08, 2019

**Expiration Date: Feb 08, 2027**

Certification performed in accordance with "EPA Traceability Protocol for Assay and Certification of Gaseous Calibration Standards (May 2012)" document EPA 600/R-12/531, using the assay procedures listed. Analytical Methodology does not require correction for analytical interference. This cylinder has a total analytical uncertainty as stated below with a confidence level of 95%. There are no significant impurities which affect the use of this calibration mixture. All concentrations are on a volume/volume basis unless otherwise noted.

Do Not Use This Cylinder below 100 psig, i.e. 0.7 megapascals.

**ANALYTICAL RESULTS**

Component	Requested Concentration	Actual Concentration	Protocol Method	Total Relative Uncertainty	Assay Dates
NOX	50.00 PPM	50.76 PPM	G1	+/- 1.4% NIST Traceable	02/01/2019, 02/08/2019
NITRIC OXIDE	50.00 PPM	50.76 PPM	G1	+/- 1.4% NIST Traceable	02/01/2019, 02/08/2019
NITROGEN	Balance			-	

**CALIBRATION STANDARDS**

Type	Lot ID	Cylinder No	Concentration	Uncertainty	Expiration Date
NTRM	1606608	CC442565	50.42 PPM NITRIC OXIDE/NITROGEN	+/- 0.8%	Jun 27, 2020
PRM	12367	APEX1099237	10.0 PPM NITROGEN DIOXIDE/AIR	+/- 1.5%	Jun 02, 2017
GMIS	1114201605	CC506716	4.995 PPM NITROGEN DIOXIDE/NITROGEN	+/- 2.0%	Nov 14, 2019

The SRM, PRM or RGM noted above is only in reference to the GMIS used in the assay and not part of the analysis.

**ANALYTICAL EQUIPMENT**

Instrument/Make/Model	Analytical Principle	Last Multipoint Calibration
Nicolet 6700 AMP0900100	FTIR	Jan 28, 2019
Nicolet 6700 AMP0900100	FTIR	Jan 28, 2019

Triad Data Available Upon Request



Signature on file

Approved for Release